
 Summary of Anomalies:
 1 - MPH Product Confidence: HDDT flag set MS 25893

Station ID	OK
MPH Product Confidence	HDDT flag set MS 25893
Flag Name	Start Stop No Pdcts Station D.Orb.
HDDT	02-APR-2000 21:47:35.990 22:17:34.670 1200 MS 25893
HDDT	02-APR-2000 22:16:36.170 22:46:34.840 1161 MS 25893
SPH Window Information	OK
Command Word Echo Summary	OK
Instrument Status 1A	OK
Instrument Status 1B	OK
Instrument Status 2	OK
Integration Times Channel 1	OK
Co-Adding and Cluster Mode Flags	OK
Integration Times Band 2A	OK
Integration Times Band 2B	OK
Integration Times Band 3	OK
Integration Times Band 4	OK
Scan Mirror Position	>>GOME polar viewline GMNNPT35 operated as planned (between 10/03/2000 - 03/04/2000)
Polarisation Detectors	OK
FPA Temperatures A	OK
FPA Temperatures B	OK
Charge Amp Temperatures	OK
Other Temperatures A	OK
DDHU Temperatures	OK
Optical Bench Temperatures	OK
Other Temperatures B	OK
Calibr. Lamp and Instr. Status 3	OK
Scan Mirror Motor Current	OK
Selected Temperature A	OK
Selected Temperature B	OK
Selected Temperature C	OK
Channel 1 Summation	OK
Channel 2 Summation	OK
Channel 3 Summation	OK
Channel 4 Summation	OK
Log pages	OK
331/318 nm Uncal. Line Ratio	OK
Uncal. PMDs as RGB signal	OK
780 nm Uncal. Intensity	OK

 GOME Memory Dump Check

2000

GOME Daily Dump Check of processing day 000402

8916 differences found in RAM1 area
 0 differences found in PROMCOPY area
 5 differences found in E2PROMCOPY area

Address	Actual	Template
12FDE	0	FF
12FDF	7	FF
13003	0	1
13007	0	1
1300B	1	10

0 differences found in E2PROM area
 0 differences found in RAM2 area

0 differences found in PROM area