
Summary of Anomalies:

- 1 - solar calibration not performed,
 but lamp calibration nominally from 10:38:56 - 10:40:53
 shutter was not opened

- 2 - KS, dump Orb. 31975 EGOI missing
 19:08:56 - 20:48:31

- 3 - instrument in nadir static view dump
 Orb. 31973, 17:25:39 - 18:50:20

```

Station ID                    >>KS, dump Orb. 31975 EGOI missing
                              19:08:56 - 20:48:31
MPH Product Confidence        OK
SPH Window Information        OK
Command Word Echo Summary    OK
Instrument Status 1A          OK
Instrument Status 1B          >>cover remains closed all day
                              no sun calibration performed
Instrument Status 2            OK
Integration Times Channel 1    OK
Co-Adding and Cluster Mode Flags OK
Integration Times Band 2A      OK
Integration Times Band 2B      OK
Integration Times Band 3       OK
Integration Times Band 4       OK
Scan Mirror Position          >>instrument in nadir static view dump
                              Orb. 31973, 17:25:39 - 18:50:20
Polarisation Detectors        OK
FPA Temperatures A            OK
FPA Temperatures B            OK
Charge Amp Temperatures       OK
Other Temperatures A          OK
DDHU Temperatures             OK
Optical Bench Temperatures    OK
Other Temperatures B          OK
Calibr. Lamp and Instr. Status 3 OK
Scan Mirror Motor Current     OK
Selected Temperature A        OK
Selected Temperature B        OK
Selected Temperature C        OK
Channel 1 Summation            OK
Channel 2 Summation            OK
Channel 3 Summation            OK
Channel 4 Summation            OK
Log pages                      OK
331/318 nm Uncal. Line Ratio   OK
Uncal. PMDs as RGB signal     OK
780 nm Uncal. Intensity       OK
  
```

8933 differences found in RAM1 area
 0 differences found in PROMCOPY area
 6 differences found in E2PROMCOPY area

Address	Actual	Template
10F45	E7	A7
12FDE	0	FF
12FDF	7	FF

```
13003 0 1
13007 0 1
1300B 1 10
0 differences found in E2PROM area
0 differences found in RAM2 area
0 differences found in PROM area
```