```
GOME Daily Reports Analysis
                                    17 Sep
Station ID
                           OK
MPH Product Confidence
                                 OK
SPH Window Information
                                 OK
                                 OK
Command Word Echo Summary
Instrument Status 1A
                                 OK
Instrument Status 1B
                                 OK
Instrument Status 2
                                 OK
Integration Times Channel 1
                                 OK
Co-Adding and Cluster Mode Flags
                                 OK
Integration Times Band 2A
                                 OK
Integration Times Band 2B
                                 OK
Integration Times Band 3
                                 OK
Integration Times Band 4
                                 OK
Scan Mirror Position
                                  >>GOME polar view timeline GMNSPT31 as planned
Polarisation Detectors
FPA Temperatures A
                                 OK
FPA Temperatures B
                                 OK
Charge Amp Temperatures
                                 OK
Other Temperatures A
                                 OK
DDHU Temperatures
                                 OK
Optical Bench Temperatures
                                 OK
Other Temperatures B
                                 OK
Calibr. Lamp and Instr. Status 3
                                 >> Solar calibration performed
                                      in dump Orb. 33515
                                  start of sun calibration 10:45:44
                                  intensity of sun calibration:
                                  ~18000 BU ->PMD2 readouts analysed with ERGO;
                                  (note: due to seasonal effects nominal values
                                  range between ~16000 - ~19000 BU)
                                 OK
Scan Mirror Motor Current
Selected Temperature A
                                 OK
Selected Temperature B
                                 OK
Selected Temperature C
                                 OK
Channel 1 Summation
                                 OK
Channel 2 Summation
                                 OK
Channel 3 Summation
                                 OK
Channel 4 Summation
                                 OK
Log pages
                                 OK
331/318 nm Uncal. Line Ratio
                                 OK
Uncal. PMDs as RGB signal
                                 OK
780 nm Uncal. Intensity
                                 OK
    GOME Memory Dump Check
                                                  17 Sep
                                                                     2001
        8830 differences found in RAM1 area
           0 differences found in PROMCOPY area
          5 differences found in E2PROMCOPY area
Address Actual Template
12FDE
                   FF
           0
 12FDF
           7
                   FF
 13003
           0
                    1
 13007
           0
                    1
 1300B
           1
                   10
           0 differences found in E2PROM area
           0 differences found in RAM2 area
```

0 differences found in PROM area

