```
*********************************
Summary of Anomalies:
station info
GS dump orbit 37547 EGOI data gap 02:39:00 - 02:47:46
GS dump orbit 37548 EGOI data gap 02:58:07 - 04:21:03
*******************************
______
  GOME Daily Reports Analysis 26 Jun
Station ID
                         >>GS dump orbit 37547 EGOI data gap 02:39:00 - 02:47:46
                                GS dump orbit 37548 EGOI data gap 02:58:07 - 04:21:03
MPH Product Confidence
SPH Window Information
                             OK
Command Word Echo Summary
                             OK
Instrument Status 1A
                             OK
Instrument Status 1B
                             OK
Instrument Status 2
                             OK
Integration Times Channel 1
                             OK
Co-Adding and Cluster Mode Flags OK
Integration Times Band 2A
                             OK
Integration Times Band 2B
                             OK
Integration Times Band 3
                             OK
Integration Times Band 4
                             OK
Scan Mirror Position
                             OK
Polarisation Detectors
                             OK
FPA Temperatures A
                             OK
FPA Temperatures B
                             OK
Charge Amp Temperatures
                             OK
Other Temperatures A
                             OK
DDHU Temperatures
                             OK
Optical Bench Temperatures
                             OK
Other Temperatures B
                             OK
Calibr. Lamp and Instr. Status 3 >> Solar calibration performed
                                 in dump Orb. 37553
                              start of sun calibration 12:57:05.30
                              intensity of sun calibration:
                              ~16650 BU ->PMD2 readouts analysed with ERGO;
                              (note: due to seasonal effects nominal values
                              range between ~16000 - ~19000 BU
Scan Mirror Motor Current
                             OK
Selected Temperature A
                             OK
Selected Temperature B
                             OK
Selected Temperature C
                             OK
Channel 1 Summation
                             OK
Channel 2 Summation
                             OK
Channel 4 Summation
                             OK
Log pages
                             OK
331/318 nm Uncal. Line Ratio
                             OK
Uncal. PMDs as RGB signal
780 nm Uncal. Intensity
                             OK
   GOME Memory Dump Check
                                    26 Jun
-----
       8809 differences found in RAM1 area
         0 differences found in PROMCOPY area
         6 differences found in E2PROMCOPY area
Address Actual Template
12FDE
          0
                 FF
12FDF
                 FF
13003
          0
                 1
13007
          0
                 1
 1300B
                 10
```

1



17FFF

FF

0 differences found in E2PROM area 0 differences found in RAM2 area

0 differences found in PROM area

