
Summary of Anomalies:

station info

BE orbit 60449 EGOI data missing 01:45:57 - 01:56:45
 BE orbit 60451 EGOI data missing 05:05:57 - 05:12:46
 MM orbit 60454 EGOI data missing 10:48:00 - 10:59:42
 MM orbit 60455 EGOI data missing 12:27:58 - 12:40:31
 MA orbit 60455 EGOI data missing 11:48:56 - 11:54:34
 BE orbit 60456 EGOI data missing 13:02:49 - 13:14:36
 BE orbit 60457 EGOI data missing 14:41:20 - 14:54:19
 KS orbit 60457 EGOI data missing 14:56:18 - 15:07:49
 MM orbit 60459 EGOI data missing 19:05:30 - 19:18:09
 MA orbit 60460 EGOI data missing 19:44:18 - 19:56:44
 MA orbit 60461 EGOI data missing 21:23:11 - 21:36:21
 MI orbit 60457 EGOI data gap 15:24:02 - 15:26:25

instrument info

EGOI

1 - complete solar calibration measurements available
 start time 11:43:03.820 , orbit 60455 (4th KS orbit),
 (increase of intensity of PMD readouts during available
 solar calibration measurements data:
 15973 BU ->PMD2 readouts analysed with ERGO.

 GOME Daily Reports Analysis 12 Nov 2006

| | |
|----------------------------------|-----------|
| Station ID | see above |
| MPH Product Confidence | OK |
| SPH Window Information | OK |
| Command Word Echo Summary | OK |
| Instrument Status 1A | OK |
| Instrument Status 1B | OK |
| Instrument Status 2 | OK |
| Integration Times Channel 1 | OK |
| Co-Adding and Cluster Mode Flags | OK |
| Integration Times Band 2A | OK |
| Integration Times Band 2B | OK |
| Integration Times Band 3 | OK |
| Integration Times Band 4 | OK |
| Scan Mirror Position | OK |
| Polarisation Detectors | OK |
| FPA Temperatures A | OK |
| FPA Temperatures B | OK |
| Charge Amp Temperatures | OK |
| Other Temperatures A | OK |
| DDHU Temperatures | OK |
| Optical Bench Temperatures | OK |
| Other Temperatures B | OK |
| Calibr. Lamp and Instr. Status 3 | OK |
| Scan Mirror Motor Current | OK |
| Selected Temperature A | OK |
| Selected Temperature B | OK |
| Selected Temperature C | OK |
| Channel 1 Summation | OK |
| Channel 2 Summation | OK |
| Channel 4 Summation | OK |
| Log pages | OK |
| 331/318 nm Uncal. Line Ratio | OK |
| Uncal. PMDs as RGB signal | OK |
| 780 nm Uncal. Intensity | OK |