
Summary of Anomalies:

station info

KS orbit 62266 EGOI data missing	18-MAR-2007 23:57:17.095 - 19-MAR-2007 00:03:03.081	345.98600 [sec]
BE orbit 62267 EGOI data missing	19-MAR-2007 01:54:15.448 - 19-MAR-2007 02:05:40.702	685.25400 [sec]
BE orbit 62269 EGOI data missing	19-MAR-2007 05:15:01.157 - 19-MAR-2007 05:20:24.638	323.48100 [sec]
MM orbit 62269 EGOI data missing	19-MAR-2007 05:54:49.144 - 19-MAR-2007 06:00:46.349	357.20500 [sec]
BE orbit 62274 EGOI data missing	19-MAR-2007 13:11:03.647 - 19-MAR-2007 13:23:14.564	730.91700 [sec]
BE orbit 62275 EGOI data missing	19-MAR-2007 14:50:02.054 - 19-MAR-2007 15:02:44.918	762.86400 [sec]
MM orbit 62278 EGOI data missing	19-MAR-2007 20:53:28.607 - 19-MAR-2007 21:06:12.212	763.60500 [sec]
MM orbit 62279 EGOI data missing	19-MAR-2007 22:33:37.762 - 19-MAR-2007 22:46:00.193	742.43100 [sec]
MA orbit 62279 EGOI data missing	19-MAR-2007 21:31:52.259 - 19-MAR-2007 21:44:50.579	778.32000 [sec]
MI orbit 62275 EGOI data gap	19-MAR-2007 15:33:37.371 - 19-MAR-2007 15:35:10.712	93.341000 [sec]

instrument info

EGOI

1 - complete solar calibration measurements available
 start time 15:07:41.715, orbit 62275 (6th KS orbit),
 (increase of intensity of PMD readouts during available
 solar calibration measurements data:
 15586 BU ->PMD2 readouts analysed with ERGO.

 GOME Daily Reports Analysis 19 Mar 2007

Station ID	see above
MPH Product Confidence	OK
SPH Window Information	OK
Command Word Echo Summary	OK
Instrument Status 1A	OK
Instrument Status 1B	OK
Instrument Status 2	OK
Integration Times Channel 1	OK
Co-Adding and Cluster Mode Flags	OK
Integration Times Band 2A	OK
Integration Times Band 2B	OK
Integration Times Band 3	OK
Integration Times Band 4	OK
Scan Mirror Position	>>GOME North Polar view operated
Polarisation Detectors	OK
FPA Temperatures A	OK
FPA Temperatures B	OK
Charge Amp Temperatures	OK
Other Temperatures A	OK
DDHU Temperatures	OK
Optical Bench Temperatures	OK
Other Temperatures B	OK
Calibr. Lamp and Instr. Status 3	OK
Scan Mirror Motor Current	OK
Selected Temperature A	OK
Selected Temperature B	OK
Selected Temperature C	OK
Channel 1 Summation	OK
Channel 2 Summation	OK
Channel 4 Summation	OK
Log pages	OK
331/318 nm Uncal. Line Ratio	OK
Uncal. PMDs as RGB signal	OK
780 nm Uncal. Intensity	OK