
NARROW SWATH TIMELINE GMNNOT41 executed
 continued from day 177 (2008/06/24)
 holding on until orbit 68919, time ~18:30

Summary of Anomalies:
 station info

Missing most of the orbits due to an ESRIN ground segment dissemination problem (ISS)

KS orbit 68914 EGOI data gap	25-JUN-2008 10:24:41.536 - 25-JUN-2008 10:25:52.432	70.896000 [sec]
KS orbit 68915 EGOI data gap	25-JUN-2008 12:04:08.026 - 25-JUN-2008 12:05:18.525	70.499000 [sec]
KS orbit 68916 EGOI data gap	25-JUN-2008 13:43:05.160 - 25-JUN-2008 13:44:17.623	72.463000 [sec]
KS orbit 68917 EGOI data gap	25-JUN-2008 15:21:17.207 - 25-JUN-2008 15:22:49.721	92.514000 [sec]
KS orbit 68918 EGOI data gap	25-JUN-2008 16:58:59.507 - 25-JUN-2008 17:00:17.306	77.799000 [sec]
KS orbit 68919 EGOI data gap	25-JUN-2008 18:37:05.776 - 25-JUN-2008 18:38:14.896	69.120000 [sec]
MA orbit 68920 EGOI data gap	25-JUN-2008 20:09:12.494 - 25-JUN-2008 20:10:33.458	80.964000 [sec]

instrument info

EGOI

- 1 - complete solar calibration measurements available
- start time 18:41:28, orbit 68919,
- (increase of intensity of PMD readouts during available
- solar calibration measurements data:
- 14450 BU ->PMD2 readouts analysed with ERGO.

 GOME Daily Reports Analysis 25 JUN 2008

Station ID	See above
MPH Product Confidence	OK
SPH Window Information	OK
Command Word Echo Summary	OK
Instrument Status 1A	OK
Instrument Status 1B	OK
Instrument Status 2	OK
Integration Times Channel 1	OK
Co-Adding and Cluster Mode Flags	OK
Integration Times Band 2A	OK
Integration Times Band 2B	OK
Integration Times Band 3	OK
Integration Times Band 4	OK
Scan Mirror Position	>> timeline GMNNOT41 executed as planned, until Orb. 68919, ~18:30
Polarisation Detectors	OK
FPA Temperatures A	OK
FPA Temperatures B	OK
Charge Amp Temperatures	OK
Other Temperatures A	OK
DDHU Temperatures	OK
Optical Bench Temperatures	OK
Other Temperatures B	OK
Calibr. Lamp and Instr. Status 3	OK
Scan Mirror Motor Current	>> pattern not repeated due to execution of timeline GMNNOT41
Selected Temperature A	OK
Selected Temperature B	OK
Selected Temperature C	OK
Channel 1 Summation	OK
Channel 2 Summation	OK
Channel 4 Summation	OK
Log pages	OK
331/318 nm Uncal. Line Ratio	OK
Uncal. PMDs as RGB signal	OK
780 nm Uncal. Intensity	OK

